Sample OxyMask Therapy Order

Therapy Order

The oxygen delivery device should be specified by a physician as part of their written order for oxygen therapy.

Flow Rate and Oxygen Concentrations

Cannula

Liter Flow	FIO2
1 L	24%
2 L	28%
Protocol	32%
4 L	36%
5 L	40%
6 L	44%

Oxymask

Liter Flow	FiO2
1 L	24% - 27%
2 L	27% - 32%
3 L	30% - 60%
4 L	33% - 65%
5 L	36% - 69%
7 L	48% - 80%
10L	53% - 85%
12L	57% - 89%
>15L	60% - 90%

Shows flow rate of O_2 therapy devices used including **OxyMask**. Here it shows a Cannula and OxyMask.

NOTE: The above liter flows to corresponding FIO2 are based on the use of one particular venturi mask utilizing [illegible] control oxygen selector. Always read the information with any type of [illegible] et up according to manufacturer's specifications.

"The oxygen delivery device should be specified by a physician as part of their written order for oxygen therapy."

OxyMask